

Lovibond® Water Testing

Tintometer® Group



Comparator 2000+

More than 400 different test discs available



- More than 400 different test discs available
- Compensation for coloured and turbid samples
- Guaranteed constancy of the coloured glass standards
- Integrated prism

Part Number: 142000

Usable in different applications

The Comparator 2000+ is versatile: it is used in swimming pools as well as in research centers, universities or general drinking water treatment.

Daylight lamp

So that you can always measure safely and stably under unfavorable visibility conditions, we recommend the use of the (optional) mains or battery-powered daylight lamp. So the lighting conditions always remain the same. No matter where you are measuring.

Various analyses possible

For the Comparator 2000+, more than 400 color discs and numerous reagents are available. Your combination is guaranteed. And if you have to change them once, it works without opening the comparator.

Integrated prism

The integrated prism visually brings the glass standards of the color disc and the colored sample into view.

Industry

Chemical Industry | Food and Beverage Industry | Industries Others | Municipalities | NGO | Oil and Gas | Pharmaceutical Industry | Pool Public | Pool Services | Power and Energy

Application

Beer, Malts and Caramels | Boiler Water | Chemicals | Coatings | Cooling Water | Disinfection Control | Drinking Water Treatment | Edible Oils and Fats | Food and Beverage | Galvanization | Petroleum Oils and Waxes | Pool Water Control | Pool Water Treatment | Waste-Water Treatment

Comparator 2000+

The Comparator 2000+ is a versatile colorimetric system for water analysis. Comfortable handling, no compromise on accuracy and reproducibility: The Comparator 2000+ can be used for many applications. Swimming pools, research or drinking water treatment are just a few examples.

Measuring Range

Test Name	Measuring Range	Chemical Method
	0 - 30 mg/L Pt	(APHA) Platinum Cobalt Standard Method
	10 - 90 mg/L Pt	(APHA) Platinum Cobalt Standard Method
	30 - 70 mg/L Pt	(APHA) Platinum Cobalt Standard Method
	50 - 500 mg/L Pt	(APHA) Platinum Cobalt Standard Method
	70 - 250 mg/L Pt	(APHA) Platinum Cobalt Standard Method
Aluminium T	0 - 0.5 mg/L Al	Eriochrom Cyanine R
Amine	0 - 2 mg/L NH ₃	Methyl Orange
Amine T	1.0 - 10 mg/L NH ₃	Bromocresol Green
Ammonia Nes. L	0.02 - 0.2 mg/L N	
Ammonia Nes. L	0.2 - 0.52 mg/L N	
Ammonia Nes. L	0.56 - 1.2 mg/L N	
Ammonia Nes. L	1.2 - 2.0 mg/L N	
Ammonia T	0 - 0.4 mg/L N	Indophenole Blue
Ammonia T	0 - 1.0 mg/L N	Indophenole Blue
Ammonia T	0 - 10 mg/L N	Indophenole Blue
Bromine T	0.2 - 2.0 mg/L Br ₂	DPD
Bromine T	0.5 - 6.0 mg/L Br ₂	DPD
Bromine T	1.0 - 10 mg/L Br ₂	DPD
Chlorine dioxide T	0.04 - 0.57 mg/l ClO ₂	DPD
Chlorine dioxide T	0.19 - 1.9 mg/l ClO ₂	DPD
Chlorine dioxide T	0.25 - 5.0 mg/l ClO ₂	Potassium Iodide
Chlorine dioxide T	0.38 - 1.52 mg/l ClO ₂	DPD
Chlorine HR T	2.0 - 10 mg/L Cl ₂	Potassium Iodide
Chlorine HR T	5.0 - 50 mg/L Cl ₂	Potassium Iodide
Chlorine HR T	5.0 - 250 mg/L Cl ₂	Potassium Iodide
Chlorine L	0.01 - 0.1 mg/L Cl ₂	DPD
Chlorine L	0.1 - 1.0 mg/L Cl ₂	DPD
Chlorine L	0.1 - 1.0 mg/L Cl ₂	DPD
Chlorine L	0.1 - 2.0 mg/L Cl ₂	DPD
Chlorine L	0.02 - 0.2 mg/L Cl ₂	DPD
Chlorine L	0.02 - 0.3 mg/L Cl ₂	DPD
Chlorine L	0.2 - 0.8 mg/L Cl ₂	DPD
Chlorine L	0.2 - 4.0 mg/L Cl ₂	DPD
Chlorine L	0.05 - 0.5 mg/L Cl ₂	DPD
Chlorine L	1.0 - 4.0 mg/L Cl ₂	DPD
Chlorine T	0.01 - 0.1 mg/L Cl ₂	DPD
Chlorine T	0.1 - 1.0 mg/L Cl ₂	DPD
Chlorine T	0.1 - 1.0 mg/L Cl ₂	DPD
Chlorine T	0.1 - 1.0 mg/L Cl ₂	DPD
Chlorine T	0.1 - 1.0 mg/L Cl ₂	DPD
Chlorine T	0.1 - 2.0 mg/L Cl ₂	DPD
Chlorine T	0.1 mg/L Cl ₂	o-Tolidine
Chlorine T	0.02 - 0.2 mg/L Cl ₂	DPD
Chlorine T	0.02 - 0.3 mg/L Cl ₂	DPD
Chlorine T	0.02 - 0.5 mg/L Cl ₂	DPD
Chlorine T	0.2 - 0.8 mg/L Cl ₂	DPD
Chlorine T	0.2 - 4.0 mg/L Cl ₂	DPD
Chlorine T	0.05 - 0.5 mg/L Cl ₂	DPD
Chlorine T	0.5 - 4.0 mg/L Cl ₂	DPD
Chlorine T	0.5 - 6.0 mg/L Cl ₂	DPD
Chlorine T	0.15 - 2.0 mg/L Cl ₂	o-Tolidine
Chlorine T	1.0 - 4.0 mg/L Cl ₂	DPD
Chlorine T	1.0 - 5.0 mg/L Cl ₂	o-Tolidine

Test Name	Measuring Range	Chemical Method
Chlorine T	1.5 mg/L Cl ₂	DPD
Chlorine T	2.0 - 5.0 mg/L Cl ₂	DPD
Chlorine T	2.0 - 10.0 mg/L Cl ₂	DPD
Chlorine T	2.0 - 10 mg/L Cl ₂	DPD
Chromium L	0.4 - 4.0 mg/L Cr	Diphenylcarbazide
Copper HR T	0 - 4.0 mg/L Cu	Zincon
Copper LR T	0 mg/L Cu	Zincon
DEHA L	0.016 - 0.16 mg/L DE-HA	PPST
Fluoride T	0 - 1.6 mg/L F ⁻	SPADNS
H ₂ O ₂ LR T	0.1 mg/L H ₂ O ₂	DPD / Catalyst
H ₂ O ₂ LR T	0.01 mg/L H ₂ O ₂	DPD / Catalyst
H ₂ O ₂ LR T	0.05 - 0.5 mg/L H ₂ O ₂	DPD / Catalyst
Hardness total P	0 - 60 mg/L CaCO ₃	Murexide
Hydrazine L	0 - 0.4 mg/L N ₂ H ₄	PDMAB
Hydrazine L	0 - 1.0 mg/L N ₂ H ₄	PDMAB
Hydrazine P	0 - 0.5 mg/L N ₂ H ₄	PDMAB
Hydrazine P	0.02 - 0.2 mg/L N ₂ H ₄	PDMAB
Iodine T	0.4 - 3.6 mg/L I	DPD
Iodine T	0.7 - 14 mg/L I	DPD
Iron	0.01 - 0.1 mg/L Fe	PPST
Iron HR T	1.0 - 10 mg/L Fe	PPST
Iron LR T	0.1 - 1.0 mg/L Fe	PPST
Iron LR T	0.02 - 0.3 mg/L Fe	PPST
Mn T	0 - 4.0 mg/L Mn	Formaloxime
Molybdate HR T	5.0 - 50 mg/L MoO ₄	Thioglycolate
Molybdate HR T	10 - 150 mg/L MoO ₄	Thioglycolate
Molybdate L P	0 - 10 mg/L MoO ₄	TRION
Nitrate T	0.1 - 1.0 mg/L N	Zinc Reduction / NED
Nitrite L	0.001 - 0.02 mg/L N	N-(1-Naphthyl)-ethylen-diamine
Nitrite LR T	0.002 - 0.05 mg/L N	N-(1-Naphthyl)-ethylen-diamine
Nitrite LR T	0.05 - 0.5 mg/L N	N-(1-Naphthyl)-ethylen-diamine
Oxygen dissolved L	2.0 - 12 mg/L O ₂	
Ozone T	0 - 0.5 mg/L O ₃	Potassium Indigothiosul-fonate
Ozone T	0.01 - 0.1 mg/L O ₃	DPD
Ozone T	0.1 - 1.0 mg/L O ₃	DPD
Ozone T	0.05 - 0.45 mg/L O ₃	DPD
Phosphate HR T	0 - 40 mg/L P	Vanadomolybdate
Phosphate HR T	0 - 100 mg/L P	Vanadomolybdate
Phosphate L	0 - 80 mg/L P	Ammonium Molybdate / Hydroquinone
Phosphate L	10 - 100 mg/L P	Vanadomolybdate
Phosphate LR T	0 - 4.0 mg/L P	Phosphomolybdenum Blue
Phosphate L T	0.2 - 2 mg/L PO ₄	Ammonium Molybdate / Potassium Antimonyl Tartrate / Ascorbic Acid
pH value L	4.0 - 11 pH	Universal Mixing Indica-tor
pH value L	6.0 pH	Bromothymol Blue
pH value L	8.0 - 9.6 pH	Thymol Blue
pH value T	1.0 - 2.6 pH	Cresol Purple
pH value T	1.2 - 2.8 pH	Thymol Blue
pH value T	2.8 - 4.4 pH	Bromophenol Blue
pH value T	3.6 pH	Bromocresol Green
pH value T	4.4 - 6.0 pH	Methyl Red
pH value T	5.2 - 6.8 pH	Bromocresol Purple
pH value T	6.0 - 7.6 pH	Bromothymol Blue

Test Name	Measuring Range	Chemical Method
pH value T	6.8 - 8.4 pH	Phenol Red
pH value T	7.0 - 8.0 pH	Phenol Red
pH value T	7.2 - 8.8 pH	Cresol Red
pH value T	7.6 - 9.2 pH	Cresol Purple
pH value T	8.0 - 9.6 pH	Thymol Blue
QAC HR T	0 - 200 mg/L QAC	Sulphoneptaleine
QAC LR T	0 - 20 mg/L QAC	Sulphoneptaleine
Silica L	0.1 - 1.0 mg/L SiO ₂	Silicomolybdate
Silica L	0.2 - 1.0 mg/L SiO ₂	Silicomolybdate
Silica L	1.0 - 20 mg/L SiO ₂	Silicomolybdate
Silica L	2.5 - 25 mg/L SiO ₂	Silicomolybdate
Silica T	0.4 - 4.0 mg/L SiO ₂	Silicomolybdate
Silica T	1.0 - 10 mg/L SiO ₂	Silicomolybdate
Sodium hypochlorite T	2 - 16 % NaOCl	Potassium Iodide
Sugar	0 - 100	
Sulphide T	0 - 0.5 mg/L S ²⁻	DPD / Catalyst
Zinc HR T	0 mg/L Zn	Zincon / EDTA
Zinc LR T	0 mg/L Zn	Zincon / EDTA

Accessories

Title	Part Number
Bulb for Daylight cabinet 2000	124100
Daylight cabinet, battery operated	142050
Diffuser plate for cell holder	142070
Plastic cell, 13,5 mm path length, with grey lid, set of 10	145500
Plastic cell, 13,5 mm path length, with grey lid, set of 5	145505
Plastic cell, 13,5 mm path length, with grey lid, set of 100	145510
Water sampler, AF 631	170500
Daylight cabinet, mains operated	171010
External power supply	182049
Lamp 2.5 V, 200 mA for Daylight cabinet TK 102 (change in pairs)	332279
Extraction tube AF 260	352600
Glass cell, 13,5 mm path length, with white lid, DB 424/S, set of 5	354243
Stopper for glass cell DB 424/S, white	354250
Stopper for W 680/OG/40 glass cell, 40 mm path length	354270
Glas stirring rod, 20 cm length	364080
Stirring rod, 13 cm length	364100
Stirring rod, 10 cm length	364109
Glas stirring rod, 10 cm length	364110
Stirring rod, 13 cm length, set of 10	364120
Stirring rod, 10 cm length, set of 10	364130
Nitrate test tube	366220
500 ml bottle, AF 631	375069
Cleaning brush, 20 cm length	380030
Brush, 11 cm length	380230
Measuring spoon, 1 g	384930
Dilution set sodium hypochlorite	414470
Cuvette stand for 10 round cuvettes Ø 16 mm	418957
Plastic funnel with handle (white)	471007
W 680/OG/2,5 glass cell, 2,5 mm path length	606780
W 680/OG/5 glass cell, 5 mm path length	606790
W 680/OG/13 glass cell, 13 mm path length	606820
W 680/OG/13,5 glass cell, 13,5 mm path length	606830
W 680/OG/25 glass cell, 25 mm path length	606860
W 680/OG/26 glass cell, 26 mm path length	606870
W 680/OG/40 glass cell, 40 mm path length	606890

Tintometer GmbH

Lovibond® Water Testing
Schleefstraße 8-12
44287 Dortmund
Tel.: +49 (0)231/94510-0
Fax: +49 (0)231/94510-30
sales@lovibond.com
www.lovibond.com
Germany

The Tintometer Limited

Lovibond House
Sun Rise Way
Amesbury, SP4 7GR
Tel.: +44 (0)1980 664800
Fax: +44 (0)1980 625412
sales@lovibond.uk
www.lovibond.com
UK

Tintometer China

Room 1001, China Life Tower
16 Chaoyangmenwai Avenue,
Beijing, 100020
Tel.: +86 10 85251111 App. 330
Fax: +86 10 85251001
chinaoffice@tintometer.com
www.lovibond.com
China

Tintometer South East Asia

Unit B-3-12, BBT One Boulevard,
Lebuh Nilam 2, Bandar Bukit Tinggi,
Klang, 41200, Selangor D.E
Tel.: +60 (0)3 3325 2285/6
Fax: +60 (0)3 3325 2287
lovibond.asia@tintometer.com
www.lovibond.com
Malaysia

Tintometer Brazil

Caixa Postal: 271
CEP: 13201-970
Jundiaí – SP
Tel.: +55 (11) 3230-6410
sales@tintometer.com.br
www.lovibond.com.br
Brazil

Tintometer Inc.

6456 Parkland Drive
Sarasota, FL 34243
Tel: 941.756.6410
Fax: 941.727.9654
sales@lovibond.us
www.lovibond.com
USA

Tintometer India Pvt. Ltd.

Door No: 7-2-C-14, 2nd, 3rd & 4th Floor
Sanathnagar Industrial Estate,
Hyderabad, 500018
Telangana
Tel: +91 (0) 40 23883300
Toll Free: 1 800 599 3891/ 3892
indiaoffice@lovibond.in
www.lovibondwater.in
India

Tintometer Spain

Postbox: 24047
08080 Barcelona
Tel.: +34 661 606 770
sales@tintometer.es
www.lovibond.com
Spain

Technical changes without notice
Printed in Germany

Lovibond® and Tintometer® are Trademarks of the Tintometer Group of Companies